Physics

Chemistry · Biology


Technology



Lehr- und Didaktiksysteme LD Didactic GmbH

Leyboldstrasse 1 · D-50354 Huerth

06/05-W97-Iv/Sel



Safety notes

When the electron deflection tube is operated at high voltages over 5 kV, X-rays are generated, whose dose rate is below the permissible limit according to the German X-ray regulations if the current-limited high-voltage power supplies 10 kV (521 70) referred to in this instruction sheet are used.

The connection of the electron deflection tube with grounded anode given in this instruction sheet requires a high-voltage enduring voltage source for the cathode heating.

- Do not operate the electron deflection tube with high voltages over 5 kV.
- Use no other power supplies than high-voltage power supplies 10 kV (521 70) for the anode and electrode voltages.

Danger of implosion: the electron deflection tube is a high-vacuum tube made of thin-walled glass.

- Do not expose the Electron deflection tube to mechanical stress, and connect it only if it is mounted in the tube stand.
- Treat the contact pins in the pin base with care, do not bend them, and be careful when inserting them in the tube stand.
- Treat the connections of the deflecting electrodes with care.

During operation, the Electron deflection tube is heated by the cathode heating:

 If necessary, allow the Electron deflection tube to cool down before dismounting.

The Electron deflection tube may be destroyed by voltages or currents that are too high:

- Keep to the operating parameters given in the section on technical data.

Instruction sheet 555 624

Electron deflection tube (555 624)

- 1 Connection for upper deflecting electrode
- 2 Upper deflecting electrode
- 3 Lower deflecting electrode
- 4 Connection for lower deflecting electrode
- 5 Screen
- 6 Anode
- 7 Cathode cap
- 8 Getter mirror (for maintaining the vacuum)
- 9 Pin base (for electron gun contacting)

1 Description

The electron deflection tube enables the deflection of electrons in electric and magnetic fields to be studied quantitatively and the velocity and the specific charge of the electrons to be estimated.

The electric deflection takes place in the field of two capacitor plates which are built in the tube. The magnetic deflection is achieved by the magnetic field of a pair of Helmholtz coils. The course of the ray is displayed on a screen with a cm grid.

2 Technical data

Heating voltage U_F : 6-6.5 V~ (rec.: 6.3 V~) Heating current I_F : approx. 1.5 A at 6.3 V

Anode voltage U_A : 1.5-5 kV Max. capacitor voltage: 10 kV Plate distance (geometric): 50 mm

Screen: $90 \text{ mm} \times 50 \text{ mm}$

Pressure: <10⁻⁶ hPa
Diameter: 90 mm
Total length: 270 mm
Mass: 250 g


Instruction sheet 555 624 Page 2/3


3 Putting into operation

additionally required:

1 tube stand 555 600 1 high-voltage power supply 10 kV 521 70

3.1 Mounting in the tube stand:






- Hold the electron deflection tube horizontally, and turn it so that the two pins with the greatest distance in the pin base point downwards.
- Carefully insert the pin base in the socket of the tube stand until it stops.

Pin connection:

F₁, F₂: heating filaments, C: cathode cap, A: anode

3.2 Connection to the high-voltage power supply:




- For the cathode heating connect the sockets F_1 and F_2 of the tube stand to the output on the back of the high-voltage power supply 10 kV (521 70).
- Connect the socket C of the tube stand (cathode cap) to the negative pole and the socket A (anode) to the positive pole of the 5kV/2mA output, and ground the positive pole.


To functional test:

 Slowly increase the high voltage and observe the beam in the middle of the screen that becomes increasingly brighter.

4 Experiment examples

4.1 Electric deflection





additionally required:

1 high-voltage power supply 10 kV

521 70

- Connect one capacitor plate to the positive pole of the right output and the other one to the negative pole of the left output of the second high-voltage power supply, and ground the middle sockets of the high-voltage power supply.
- Apply an accelerating voltage U_A ≤ 5 kV.
- Apply a plate voltage U_P ≤ 10 kV.

Evaluation:

The trajectory (parabola) of the electrons in the electric field E is given by


$$y = \frac{1}{2} \cdot a \cdot t^2 = \frac{1}{2} \cdot \frac{e \cdot E}{m} \cdot \frac{x^2}{v^2}$$
 with $e \cdot U_A = \frac{m}{2} \cdot v^2$ and $E = \frac{U_P}{d}$.


Here the effective plate distance d is greater than the geometric plate distance because the electric field is not homogeneous. From this

$$y = \frac{U_P}{4 \cdot U_A \cdot d} \cdot x^2$$

follows

4.2 Magnetic deflection





additionally required:

1 pair of Helmholtz coils 1 DC power supply 0 ... 16 V, 5 A 555 604 521 541

- Mount the pair of Helmholtz coils in the magnet feet, attach them at the distance marked H to the tube stand, and align them in height so that they exactly embrace the electron deflection tube.
- Apply an accelerating voltage $U_A \le 5 \text{ kV}$.
- Connect the pair of Helmholtz coils in parallel to the DC power supply and apply the current 2*I*_H.

Evaluation:

The trajectory (circle with radius r) of the electrons in the magnetic field B is given by

$$y = r - \sqrt{r^2 - x^2}$$
 with $r = \sqrt{\frac{2 \cdot m \cdot U_A}{e \cdot B}}$

because of $m \cdot \frac{v^2}{r} = e \cdot v \cdot B$ and $e \cdot U_A = \frac{m}{2} \cdot v^2$,

whereby the magnetic field of the Helmholtz coils is proportional to the current $I_{\rm H}$

$$B = \left(\frac{4}{5}\right)^{\frac{3}{2}} \cdot \frac{N \cdot I}{R}$$
 with $N = 320$ turns, $R = 6.25$ cm.